

#### **Certificate of Calibration - Wind Monitoring Station**

#### 1. Performance check of Wind Speed

Wind Sp	beed, m/s	Difference D (m/s)
Wind Speed Reading (V1)	Anemometer Value (V2)	D = V1 - V2
0.0	0.0	0.0
1.5	1.5	0.0
2.8	2.7	0.1
4.0	4.1	-0.1

#### 2. Performance check of Wind Direction

Wind Di	rection (°)	Difference D (°)
Wind Direction Reading (W1)	Marine Compass Value (W2)	$\mathbf{D} = \mathbf{W1} - \mathbf{W2}$
0	0	0.0
90	90	0.0
180	180	0.0
270	270	0.0

**Test Specification:** 

1. Performance Wind Speed Test - The wind meter was on-site calibrated against the anemometer

2. Performance Wind Direction Test - The wind meter was on-site calibrated against the marine compass at four direction





Certificate of Calibration

			Calibration	Certificati	on Informat	tion			
Cal. Date:				meter S/N:	438320	Та:	297	°К	
Operator:	Jim Tisch					Pa:	750.1	mm Hg	
Calibration	Model #:	TE-5025A	Calil	brator S/N:	3864				
								1	
		Vol. Init	Vol. Final	ΔVol.	ΔTime	ΔΡ	ΔΗ		
	Run	(m3)	(m3)	(m3)	(min)	(mm Hg)	(in H2O)		
	1	1	2	1	1.4470	3.2	2.00		
	2	3	4	1	1.0210	6.4	4.00		
	3	5	6	1	0.9140	8.0	5.00		
	4	, 7	8	1	0.8670	8.8	5.50		
	5	9	10	1	0.7140	12.9	8.00		
			[	Data Tabula	tion			]	
			/ / Pa	V Tetd )					
	Vstd	Qstd	√ <sup>∆H</sup> (Pstd	)( <u>Tstd</u> )		Qa	√∆H( Ta/Pa )		
	(m3)	(x-axis)	y (y-ax		Va	(x-axis)	(y-axis)		
	0.9860	0.6814	1.40		0.9957	0.6881	0.8899		
	0.9818	0.9616	1.99	02	0.9915	0.9711	1.2585	1	
	0.9797	1.0719	2.22	51	0.9893	1.0824	1.4071	1	
	0.9786	1.1288	2.33	37	0.9883	1.1399	1.4757	1	
	0.9732	1.3630	2.814	46	0.9828	1.3765	1.7798		
		m=	2.065	566		m=	1.29348		
		b=	0.003	815	QA	b=	0.00199		
		r=	0.999	96		r=	0.99996		
				Calculations					
	Vstd=	ΔVol((Pa-ΔP)	)/Pstd)(Tstd/Ta	a)	Va=	ΔVol((Pa-Δ	P)/Pa)		
	Qstd=	Vstd/∆Time			Qa=				
			For subsequ	ent flow ra	te calculatio	ns:			
	Qstd=	<b>Qstd=</b> $1/m\left(\left(\sqrt{\Delta H\left(\frac{Pa}{Pstd}\right)\left(\frac{Tstd}{Ta}\right)}\right)-b\right)$				$Qa = 1/m\left(\left(\sqrt{\Delta H(Ta/Pa)}\right)-b\right)$			
	Standard	Conditions							
Tstd						RECA	LIBRATION		
Pstd	760	mm Hg						400	
A 1 1 . 1+1		Key	1120)				nnual recalibratio	-	
		ter reading (i					Regulations Part		
		eter reading perature (°K)					, Reference Meth		
		ressure (mm				1	ended Particulat		
b: intercept	the second s				tn tn	e Atmosphe	ere, 9.2.17, page	30	
m: slope									

isch Environmental, Inc. 45 South Miami Avenue illage of Cleves, OH 45002 <u>www.tisch-env.com</u> TOLL FREE: (877)263-7610 FAX: (513)467-9009

.

# CIN@TECH &

File No. MA20003/44/0009

Project No.	KTD1 - Centre	KTD1 - Centre of Excellence in Paediatrics (Children's Hospital)					
Date:	2-0	Dct-21	Next Due Date:	2-Dec-21	Operator:	SK	
Equipment No.:	A-	01-44	Model No.:	TE-5170	Serial No.	1316	
			Ambient Conditi	on			
Temperatu	re, Ta (K)	303	Pressure, Pa (mmH	[g)	758.1		

Orifice Transfer Standard Information							
Serial No.	3864	Slope, mc	0.05846	Intercept, bc	-0.00313		
Last Calibration Date:	11-Jan-21	mc x Qstd + bc = $[\Delta H x (Pa/760) x (298/Ta)]^{1/2}$					
Next Calibration Date:	11-Jan-22	Qstd = { $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ -bc} / mc					

Calibration of TSP Sampler								
Calibration		Orfice	HVS					
Point	ΔH (orifice), in. of water	$[\Delta H \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}$	Qstd (CFM) X - axis	$\Delta W$ (HVS), in. of water	$[\Delta W \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}$ Y-axis			
1	13.6	3.65	62.54	9.2	3.00			
2	11.2	3.31	56.75	7.2	2.66			
3	8.2	2.84	48.57	5.6	2.34			
4	5.6	2.34	40.15	3.3	1.80			
5	3.2	1.77	30.36	1.8	1.33			
Slope, mw =	ression of Y on X 0.0520		Intercept, bw	-0.251	5			
	coefficient* =	0.9978	_					
*If Correlation (	Coefficient < 0.99	0, check and recalibrate.						
		Set Point C	Calculation					
From the TSP Fi	ield Calibration C	urve, take Qstd = 43 CFM						
From the Regres	sion Equation, the	e "Y" value according to						
Therefore, Se	et Point; W = ( my	$\mathbf{mw} \mathbf{x} \mathbf{Qstd} + \mathbf{bw} = [\Delta \mathbf{W}]$ v x Qstd + bw ) <sup>2</sup> x ( 760 / Pa ) x (		98/Ta)] <sup>1/2</sup> 4.01				
Remarks:								
Conducted by:	Wong Shi	ng Kwai Signature	<u> </u>	<u>Д.                                    </u>	Date: 2-Oct-21			
Checked by:	Henry 1	Leung Signature	: \-lem	, Xoy	Date: 2-Oct-21			

# CIN@TECH 4

File No. MA20003/04/0008

Project No.	KER 1 - Future	e Residential De	velopment at Kerry Godow	vn			
Date:	2-0	Det-21	Next Due Date:	2-Dec-21	Operator:	SK	
Equipment No.:	A-01-04		Model No.:	TE 5170	Serial No.	10595	
			Ambient Conditi	on			
Temperatu	ure, Ta (K)	303	Pressure, Pa (mmH	łg)	758.1		

Orifice Transfer Standard Information							
Serial No.	3864	Slope, mc	0.05846	Intercept, bc	-0.00313		
Last Calibration Date:	11-Jan-21	mc x Qstd + bc = $[\Delta H x (Pa/760) x (298/Ta)]^{1/2}$					
Next Calibration Date:	11-Jan-22	Qstd = { $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ -bc} / mc					

Calibration of TSP Sampler								
Calibration		Orfice	HVS					
Point	ΔH (orifice), in. of water	$[\Delta H x (Pa/760) x (298/Ta)]^{1/2}$	Qstd (CFM) X - axis	$\Delta W$ (HVS), in. of water	$\frac{[\Delta W \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}}{Y-axis}$			
1	13.4	3.63	62.07	9.1	2.99			
2	10.6	3.22	55.22	7.0	2.62			
3	8.4	2.87	49.16	5.6	2.34			
4	5.2	2.26	38.69	3.2	1.77			
5	3.0	1.72	29.40	2.1	1.44			
Slope, mw =	ression of Y on X 0.0482 coefficient* =		Intercept, bw	-0.028	80			
*If Correlation C	Coefficient < 0.99	0, check and recalibrate.						
		Set Point C	alculation					
		urve, take Qstd = 43 CFM						
	-	e "Y" value according to <b>mw x Qstd + bw = [<math>\Delta W</math></b> v x Qstd + bw ) <sup>2</sup> x ( 760 / Pa ) x (		98/Ta)] <sup>1/2</sup> 4.26				
Remarks:								
Conducted by:	Wong Shi	ng Kwai Signature	: <u>k</u>	<u>у</u>	Date: 2-Oct-21			
Checked by:	Henry	Leung Signature	:_ \-len	N- - Nor	Date: 2-Oct-21			



File No. MA20003/41/0008

Project No.	KTD 2D - Nex	t to the SOR Off	fice of Trunk Road T2 in I	Kai Tak Area			
Date:	25-Sep-21		Next Due Date:	25-Nov-21	Operator:	SK	
Equipment No.:	:		Model No.:	TE 5170	Serial No.	5280	
			Ambient Condit	tion			
Temperature, Ta (K) 302.6		Pressure, Pa (mm	Hg)	759.5			
Orifice Transfer Standard Information							

Orifice Transfer Standard Information							
Serial No.	3864	Slope, mc	0.05846	Intercept, bc	-0.00313		
Last Calibration Date:	11-Jan-21	1	mc x Qstd + bc = $[\Delta H x (Pa/760) x (298/Ta)]^{1/2}$				
Next Calibration Date:	11-Jan-22	Qstd = { $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ -bc} / mc					

Calibration of TSP Sampler								
Calibration		Orfice			HVS			
Point	$\Delta H$ (orifice), in. of water	[ΔH x (Pa/760) x (298/Ta)] <sup>1/2</sup>	Qstd (CFM) X - axis	$\Delta W$ (HVS), in. of water	$[\Delta W \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}$ Y-axis			
1	13.4	3.63	62.17	8.9	2.96			
2	11.6	3.38	57.85	7.1	2.64			
3	8.3	2.86	48.94	5.6	2.35			
4	6.0	2.43	41.62	4.0	1.98			
5	3.0	1.72	29.45	2.4	1.54			
By Linear Regression of Y on X Slope , mw =Intercept, bw :								
		urve, take Qstd = 43 CFM e "Y" value according to						
		$\mathbf{mw} \mathbf{x} \mathbf{Qstd} + \mathbf{bw} = [\Delta \mathbf{W} \mathbf{x}]$ v x Qstd + bw ) <sup>2</sup> x ( 760 / Pa ) x (						
Remarks:								
Conducted by:	Wong Shi	ng Kwai Signature:	k	火.	Date: 25-Sep-21			
Checked by:	Henry I	Leung Signature:	- le-	N. Jan J	Date: 25-Sep-21			



File No. MA20003/55/0010

Project No.	CKL 2 - Flat 10	)3 Cha Kwo Lin	g Village			
Date:	6-S	ep-21	Next Due Date:	6-Nov-21	Operator:	SK
Equipment No.:	A-(	01-55	Model No.:	TE 5170	Serial No.	1956
			Ambient Condition	on		
Temperatu	re, Ta (K)	302.7	Pressure, Pa (mmH	(g)	757.8	
				T. C		

Orifice Transfer Standard Information							
Serial No.	Serial No.         3864         Slope, mc         0.05846         Intercept, bc         -0.00313						
Last Calibration Date:	11-Jan-21	mc x Qstd + bc = $[\Delta H x (Pa/760) x (298/Ta)]^{1/2}$					
Next Calibration Date:	11-Jan-22	Qstd = { $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ -bc} / mc					

		Calibration of	TSP Sampler			
Calibration		Orfice			HVS	
Point	$\Delta H$ (orifice), in. of water	$[\Delta H \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}$	Qstd (CFM) X - axis	$\Delta W$ (HVS), in. of water		50) x (298/Ta)] <sup>1/2</sup> <b>7-axis</b>
1	12.8	3.54	60.69	9.8		3.10
2	10.8	3.26	55.75	7.6		2.73
3	8.4	2.87	49.17	6.0		2.43
4	5.6	2.34	40.16	3.6		1.88
5	3.0	1.72	29.41	1.9		1.37
By Linear Regr Slope , mw =			Intercept, bw =	-0.283	31	
Correlation of	coefficient* =	0.9979	_			
*If Correlation C	coefficient < 0.99	0, check and recalibrate.				
		Set Point C	alculation			
From the TSP Fi	eld Calibration C	urve, take Qstd = 43 CFM				
From the Regress	sion Equation, the	e "Y" value according to				
		$\mathbf{mw} \mathbf{x} \mathbf{Qstd} + \mathbf{bw} = [\Delta \mathbf{W}]$	x (Pa/760) x (29	98/Ta)] <sup>1/2</sup>		
Therefore, Se	et Point; W = ( my	$(x + y)^{2} (760 / Pa) x (760 / Pa) x ($	Ta / 298 ) =	4.40		
Remarks:						
Conducted by:	Wong Shi	ng Kwai Signature:	k	火.	Date:	6-Sep-21
Checked by:	Henry 1	Leung Signature:	- lan	J Jan J	Date:	6-Sep-21



#### File No. MA20003/18/0010

Project No.	CKL 1 - Flat 1	21 Cha Kwo Ling	g Village			
Date:	6-S	Sep-21	Next Due Date:	6-Nov-21	Operator:	SK
Equipment No.:	A-	01-18	Model No.:	TE 5170	Serial No.	0723
			Ambient Condi	tion		
Temperatu	re, Ta (K)	302.7	Pressure, Pa (mml	Hg)	757.8	

Orifice Transfer Standard Information								
Serial No.         3864         Slope, mc         0.05846         Intercept, bc         -0.00313								
Last Calibration Date:	Last Calibration Date: 11-Jan-21 $\operatorname{mc} x \operatorname{Qstd} + \operatorname{bc} = [\Delta H \times (\operatorname{Pa}/760) \times (298/\operatorname{Ta})]^{1/2}$							
Next Calibration Date:								

		Calibration of	of TSP Sampler		
Colibration		Orfice	-		HVS
Calibration Point	ΔH (orifice), in. of water	$[\Delta H x (Pa/760) x (298/Ta)]^{1/2}$	Qstd (CFM) X - axis	$\Delta W$ (HVS), in. of water	$\frac{\left[\Delta W \ge (Pa/760) \ge (298/Ta)\right]^{1/2} \ Y-axis}{axis}$
1	12.8	3.54	60.69	9.8	3.10
2	10.2	3.16	54.18	8.0	2.80
3	8.4	2.87	49.17	5.9	2.41
4	6.2	2.47	42.25	4.0	1.98
5	3.4	1.83	31.30	1.9	1.37
Slope, mw =	ression of Y on X 0.0605 coefficient* =	0.9981	Intercept, bw	-0.54	14
			—		
*If Correlation C	Coefficient < 0.99	0, check and recalibrate.	Calariation		
From the TSD Fi	iald Calibratian C	urve, take Qstd = 43 CFM	Calculation		
		-			
From the Regres	sion Equation, the	e "Y" value according to			
		mw x Qstd + bw = $[\Delta W]$	/ x (Pa/760) x (2	298/Ta)] <sup>1/2</sup>	
Therefore, Se	et Point; W = ( my	$(x + bw)^2 x (760 / Pa) x$	(Ta / 298) =	4.32	2
Remarks:					
Conducted by:	Wong Shi	ng Kwai Signaturo		<u>Х.</u>	Date: 6-Sep-21
Checked by:	Henry I	Leung Signature	:_ \-lem	J Xoz	Date: 6-Sep-21

# CIN@TECH 🤳

File No. MA20003/44/0008

Project No.	KTD1 - Centre of Excellence in Paediatrics (Children's Hospital)						
Date:	2-A	ug-21	Next Due Date:	2-C	Oct-21	Operator:	SK
Equipment No.:	A-	01-44	Model No.:	TE	-5170	Serial No	1316
			Ambient Conditi	on			
Temperatu	re, Ta (K)	302.7	Pressure, Pa (mmH	lg)		750	

Orifice Transfer Standard Information							
Serial No.         3864         Slope, mc         0.05846         Intercept, bc         -0.00313							
Last Calibration Date:	11-Jan-21	mc x Qstd + bc = $[\Delta H x (Pa/760) x (298/Ta)]^{1/2}$					
Next Calibration Date: 11-Jan-22 $Qstd = \{ [\Delta H \ x \ (Pa/760) \ x \ (298/Ta) ]^{1/2} - bc \} / mc$							

		Calibration of	TSP Sampler		
Calibration		Orfice			HVS
Point	ΔH (orifice), in. of water	$[\Delta H \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}$	Qstd (CFM) X - axis	$\Delta W$ (HVS), in. of water	$\frac{[\Delta W \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}}{Y-axis}$
1	13.8	3.66	62.69	9.2	2.99
2	11.2	3.30	56.48	7.0	2.61
3	8.3	2.84	48.63	5.4	2.29
4	6.2	2.45	42.04	3.4	1.82
5	3.2	1.76	30.21	1.8	1.32
Slope , mw = Correlation	ression of Y on X 0.0516 coefficient* = Coefficient < 0.99		Intercept, bw = _	-0.270	94
		Set Point C	alculation		
		urve, take Qstd = 43 CFM			
	-	e "Y" value according to <b>mw x Qstd + bw = [<math>\Delta W</math></b> w x Qstd + bw ) <sup>2</sup> x ( 760 / Pa ) x (		98/Ta)] <sup>1/2</sup> 	
Remarks:					
Conducted by:	Wong Shi	ng Kwai Signature	: <u>k</u>	<u>у</u>	Date: 2-Aug-21
Checked by:	Henry 1	Leung Signature	: \-lem	, Xory	Date: 2-Aug-21

# CIN@TECH 🤳

File No. MA20003/04/0007

Project No.	KER 1 - Future	e Residential Dev	velopment at Kerry Godow	√n			
Date:	2-A	Nug-21	Next Due Date:	2-Oct-21	Operator:	SK	
Equipment No.:	<u> </u>	01-04	Model No.:	TE 5170	Serial No.	10595	
			Ambient Conditi	on			
Temperatu	ure, Ta (K)	302.7	Pressure, Pa (mmH	Ig)	750		

Orifice Transfer Standard Information							
Serial No.         3864         Slope, mc         0.05846         Intercept, bc         -0.00313							
Last Calibration Date:	11-Jan-21	mc x Qstd + bc = $[\Delta H x (Pa/760) x (298/Ta)]^{1/2}$					
Next Calibration Date:	Next Calibration Date: 11-Jan-22 $Qstd = \{[\Delta H \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2} - bc\} / mc$						

Calibration of TSP Sampler					
Calibration Point	Orfice			HVS	
	ΔH (orifice), in. of water	$[\Delta H \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}$	Qstd (CFM) X - axis	$\Delta W$ (HVS), in. of water	$\frac{[\Delta W \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}}{Y-axis}$
1	13.7	3.65	62.46	9.0	2.96
2	11.4	3.33	56.98	7.0	2.61
3	8.6	2.89	49.50	5.4	2.29
4	5.2	2.25	38.50	3.2	1.76
5	3.0	1.71	29.26	2.1	1.43
By Linear Regression of Y on X Slope , mw = 0.0457 Intercept, bw = 0.0483					
Correlation coefficient* = 0.9970					
*If Correlation (	Coefficient < 0.99	0, check and recalibrate.			
Set Point Calculation					
From the TSP Field Calibration Curve, take Qstd = 43 CFM					
From the Regression Equation, the "Y" value according to					
$mw \ x \ Qstd + bw = [\Delta W \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}$ Therefore, Set Point; W = ( mw x Qstd + bw ) <sup>2</sup> x ( 760 / Pa ) x ( Ta / 298 ) =					
Remarks:					
Conducted by:	Wong Shi	ng Kwai Signature	<u> </u>	<u>у</u>	Date: 2-Aug-21
Checked by:	Henry l	Leung Signature	: \-len	, drag	Date: 2-Aug-21