

**Certificate of Calibration**

It is certified that the item under calibration has been calibrated by corresponding calibrated High Volume Sampler


Description: Digital Dust Indicator Date of Calibration 2-Oct-21  
 Manufacturer: Sibata Scientific Technology LTD. Validity of Calibration Record 2-Dec-21  
 Model No.: LD-5R  
 Serial No.: 972778  
 Equipment No.: SA-01-07 Sensitivity 0.001 mg/m3  
 High Volume Sampler No.: A-01-03 Before Sensitivity Adjustment 735 CPM  
 Tisch Calibration Orifice No.: 3864 After Sensitivity Adjustment 735 CPM


| Calibration of 1 hr TSP  |   |   |
|--|---|---|
| Calibration Point  | Laser Dust Monitor  | HVS   |
|  | Mass Concentration ( $\mu\text{g}/\text{m}^3$ )<br>X-axis | Mass concentration ( $\mu\text{g}/\text{m}^3$ )<br>Y-axis |
| 1  | 65.0  | 127.0   |
| 2  | 58.0  | 121.0   |
| 3  | 49.0  | 112.0   |
| Average  | 57.3  | 120.0   |
| By Linear Regression of Y on X<br>Slope, mw = <u>0.9404</u> Intercept, bw = <u>66.0829</u><br>Correlation coefficient* = <u>0.9991</u> |   |   |
| Set Correlation Factor   |   |   |
| Particulate Concentration by High Volume Sampler ( $\mu\text{g}/\text{m}^3$ )  |   | 120.0   |
| Particulate Concentration by Dust Meter ( $\mu\text{g}/\text{m}^3$ )   |   | 57.3  |
| Measureing time, (min)   |   | 60.0  |
| Set Correlation Factor, SCF  |   |   |
| SCF = [ K=High Volume Sampler / Dust Meter, ( $\mu\text{g}/\text{m}^3$ ) ]   |   | <u>2.1</u>  |

In-house method in according to the instruction manual:

The Dust Monitor was compared with a calibrated High Volume Sampler and The result was used to generate the Correlation Factor (CF) between the Dust Monitor and High Volume Sampler.

Those filter papers are weighted by HOKLAS laboratory (HPCT Litimed)

Calibrated by:   
 Technical Officer (Wong Shing Kwai)

Approved by:   
 Project Manager (Henry Leung)

**Certificate of Calibration**

It is certified that the item under calibration has been calibrated by corresponding calibrated High Volume Sampler


Description: Digital Dust Indicator Date of Calibration 2-Aug-21  
 Manufacturer: Sibata Scientific Technology LTD. Validity of Calibration Record 2-Oct-21  
 Model No.: LD-5R  
 Serial No.: 972781  
 Equipment No.: SA-01-10 Sensitivity 0.001 mg/m3  
 High Volume Sampler No.: A-01-03 Before Sensitivity Adjustment 734 CPM  
 Tisch Calibration Orifice No.: 3864 After Sensitivity Adjustment 734 CPM


| Calibration of 1 hr TSP   |   |   |
|---|---|---|
| Calibration Point   | Laser Dust Monitor  | HVS   |
|   | Mass Concentration ( $\mu\text{g}/\text{m}^3$ )<br>X-axis | Mass concentration ( $\mu\text{g}/\text{m}^3$ )<br>Y-axis |
| 1   | 66.0  | 131.0   |
| 2   | 57.0  | 125.0   |
| 3   | 46.0  | 116.0   |
| Average   | 56.3  | 124.0   |
| By Linear Regression of Y on X<br>Slope , mw = <u>0.7525</u> Intercept, bw = <u>81.6096</u><br>Correlation coefficient* = <u>0.9984</u> |   |   |
| Set Correlation Factor  |   |   |
| Particulate Concentration by High Volume Sampler ( $\mu\text{g}/\text{m}^3$ )   |   | 124.0   |
| Particulate Concentration by Dust Meter ( $\mu\text{g}/\text{m}^3$ )  |   | 56.3  |
| Measureing time, (min)  |   | 60.0  |
| Set Correlation Factor , SCF  |   |   |
| SCF = [ K=High Volume Sampler / Dust Meter, ( $\mu\text{g}/\text{m}^3$ ) ]  |   | <u>2.2</u>  |

In-house method in according to the instruction manual:

The Dust Monitor was compared with a calibrated High Volume Sampler and The result was used to generate the Correlation Factor (CF) between the Dust Monitor and High Volume Sampler.

Those filter papers are weighted by HOKLAS laboratory (HPCT Litimed)

Calibrated by:   
 Technical Officer (Wong Shing Kwai)

Approved by:   
 Project Manager (Henry Leung)

**Certificate of Calibration**

It is certified that the item under calibration has been calibrated by corresponding calibrated High Volume Sampler


Description: Digital Dust Indicator Date of Calibration 2-Oct-21  
 Manufacturer: Sibata Scientific Technology LTD. Validity of Calibration Record 2-Dec-21  
 Model No.: LD-5R  
 Serial No.: 972781  
 Equipment No.: SA-01-10 Sensitivity 0.001 mg/m3  
 High Volume Sampler No.: A-01-03 Before Sensitivity Adjustment 734 CPM  
 Tisch Calibration Orifice No.: 3864 After Sensitivity Adjustment 734 CPM


| Calibration of 1 hr TSP  |   |   |
|--|---|---|
| Calibration Point  | Laser Dust Monitor  | HVS   |
|  | Mass Concentration ( $\mu\text{g}/\text{m}^3$ )<br>X-axis | Mass concentration ( $\mu\text{g}/\text{m}^3$ )<br>Y-axis |
| 1  | 64.0  | 127.0   |
| 2  | 56.0  | 121.0   |
| 3  | 45.0  | 112.0   |
| Average  | 55.0  | 120.0   |
| By Linear Regression of Y on X<br>Slope, mw = <u>0.7912</u> Intercept, bw = <u>76.4835</u><br>Correlation coefficient* = <u>0.9997</u> |   |   |
| Set Correlation Factor   |   |   |
| Particulate Concentration by High Volume Sampler ( $\mu\text{g}/\text{m}^3$ )  |   | 120.0   |
| Particulate Concentration by Dust Meter ( $\mu\text{g}/\text{m}^3$ )   |   | 55.0  |
| Measureing time, (min)   |   | 60.0  |
| Set Correlation Factor, SCF  |   |   |
| SCF = [ K=High Volume Sampler / Dust Meter, ( $\mu\text{g}/\text{m}^3$ ) ]   |   | <u>2.2</u>  |

In-house method in according to the instruction manual:

The Dust Monitor was compared with a calibrated High Volume Sampler and The result was used to generate the Correlation Factor (CF) between the Dust Monitor and High Volume Sampler.

Those filter papers are weighted by HOKLAS laboratory (HPCT Litimed)

Calibrated by:   
 Technical Officer (Wong Shing Kwai)

Approved by:   
 Project Manager (Henry Leung)

# Certificate of Calibration

## Calibration Certification Information

Cal. Date: January 11, 2021      Rootsmeter S/N: 438320      Ta: 297 °K  
Operator: Jim Tisch      Pa: 750.1 mm Hg  
Calibration Model #: TE-5025A      Calibrator S/N: 3864

| Run | Vol. Init (m3) | Vol. Final (m3) | ΔVol. (m3) | ΔTime (min) | ΔP (mm Hg) | ΔH (in H2O) |
|-----|----------------|-----------------|------------|-------------|------------|-------------|
| 1   | 1              | 2               | 1          | 1.4470      | 3.2        | 2.00        |
| 2   | 3              | 4               | 1          | 1.0210      | 6.4        | 4.00        |
| 3   | 5              | 6               | 1          | 0.9140      | 8.0        | 5.00        |
| 4   | 7              | 8               | 1          | 0.8670      | 8.8        | 5.50        |
| 5   | 9              | 10              | 1          | 0.7140      | 12.9       | 8.00        |

## Data Tabulation

| Vstd (m3) | Qstd (x-axis) | $\sqrt{\Delta H \left( \frac{Pa}{Pstd} \right) \left( \frac{Tstd}{Ta} \right)}$ (y-axis) | Va     | Qa (x-axis) | $\sqrt{\Delta H \left( \frac{Ta}{Pa} \right)}$ (y-axis) |
|-----------|---------------|--|--------|-------------|---|
| 0.9860    | 0.6814        | 1.4073   | 0.9957 | 0.6881      | 0.8899  |
| 0.9818    | 0.9616        | 1.9902   | 0.9915 | 0.9711      | 1.2585  |
| 0.9797    | 1.0719        | 2.2251   | 0.9893 | 1.0824      | 1.4071  |
| 0.9786    | 1.1288        | 2.3337   | 0.9883 | 1.1399      | 1.4757  |
| 0.9732    | 1.3630        | 2.8146   | 0.9828 | 1.3765      | 1.7798  |
| QSTD      | m=            | 2.06566  | QA     | m=          | 1.29348   |
|           | b=            | 0.00315  |        | b=          | 0.00199   |
|           | r=            | 0.99996  |        | r=          | 0.99996   |

## Calculations

|  |   |     |  |
|--|---|-----|--|
| Vstd=                                  | $\Delta Vol((Pa-\Delta P)/Pstd)(Tstd/Ta)$   | Va= | $\Delta Vol((Pa-\Delta P)/Pa)$   |
| Qstd=                                  | $Vstd/\Delta Time$  | Qa= | $Va/\Delta Time$   |
| For subsequent flow rate calculations: |   |     |  |
| Qstd=                                  | $1/m \left( \left( \sqrt{\Delta H \left( \frac{Pa}{Pstd} \right) \left( \frac{Tstd}{Ta} \right)} \right) - b \right)$ | Qa= | $1/m \left( \left( \sqrt{\Delta H \left( \frac{Ta}{Pa} \right)} \right) - b \right)$ |

## Standard Conditions

Tstd: 298.15 °K

Pstd: 760 mm Hg

## Key

ΔH: calibrator manometer reading (in H2O)

ΔP: rootsmeter manometer reading (mm Hg)

Ta: actual absolute temperature (°K)

Pa: actual barometric pressure (mm Hg)

b: intercept

m: slope

## RECALIBRATION

US EPA recommends annual recalibration per 1998 40 Code of Federal Regulations Part 50 to 51, Appendix B to Part 50, Reference Method for the Determination of Suspended Particulate Matter in the Atmosphere, 9.2.17, page 30

**Certificate of Calibration - Wind Monitoring Station**

Description: Yau Lai Estate, Bik Lai House  
Manufacturer: Davis Instruments  
Model No.: Davis7440  
Serial No.: MC01010A44  
Equipment No.: SA-03-04  
Date of Calibration: 20-Aug-2021  
Next Due Date: 20-Feb-2022

## 1. Performance check of Wind Speed

| Wind Speed, m/s         |                       | Difference D (m/s) |
|-------------------------|-----------------------|--------------------|
| Wind Speed Reading (V1) | Anemometer Value (V2) | $D = V1 - V2$      |
| 0.0                     | 0.0                   | 0.0                |
| 1.5                     | 1.5                   | 0.0                |
| 2.8                     | 2.7                   | 0.1                |
| 4.0                     | 4.1                   | -0.1               |

## 2. Performance check of Wind Direction


| Wind Direction (°)          |                           | Difference D (°) |
|-----------------------------|---------------------------|------------------|
| Wind Direction Reading (W1) | Marine Compass Value (W2) | $D = W1 - W2$    |
| 0                           | 0                         | 0.0              |
| 90                          | 90                        | 0.0              |
| 180                         | 180                       | 0.0              |
| 270                         | 270                       | 0.0              |

**Test Specification:**


1. Performance Wind Speed Test - The wind meter was on-site calibrated against the anemometer

2. Performance Wind Direction Test - The wind meter was on-site calibrated against the marine compass at four direction

Calibrated by:

  
Wong Shing Kwai

Approved by:

  
Henry Leung



**APPLICANT:**      **Cinotech Consultants Limited**  
                         **RM 1710, Technology Park,**  
                         **18 On Lai Street,**

|                  |                             |
|------------------|-----------------------------|
| Test Report No.: | 00122                       |
| Date of Issue:   | 2021-05-12                  |
| Date Received:   | 2021-05-07                  |
| Test Period      | 2021-05-10 to<br>2021-05-10 |
| Next Due Date:   | 2022-05-10                  |

**ATTN:**              **Mr. Henry Leung**

|                                   |
|-----------------------------------|
| <b>Certificate of Calibration</b> |
|-----------------------------------|

**Item for calibration**

|                |                               |
|----------------|-------------------------------|
| Description    | Integrating Sound Level Meter |
| Manufacturer   | BSWA Technology               |
| Model No.      | BSWA 308                      |
| Serial No.     | 580156                        |
| Microphone No. | 580804                        |
| Equipment No.  | N-12-06                       |

**Test conditions:**

Room Temperature                      : 22-25 degree Celsius  
Relative Humidity                        : 35-70%

**Method reference:**

The Sound Level Calibrator has been calibrated in accordance with the documented procedures and using standard and instrument which are recommended by the manufacturer, or equivalent.

**Measuring equipment :**

|               |                  |
|---------------|------------------|
| Description   | Sound Calibrator |
| Manufacturer  | Brüel & Kjær     |
| Model No.     | TYPE 4231        |
| Serial No.    | 2326353          |
| Equipment No. | N-02-01          |

\*\*\*\*\*



## Test Report

### Results:

| Reference value, dB | Indication value, dB | Deviation, dB | Allowed deviation, dB |
|---------------------|----------------------|---------------|-----------------------|
| 94.0                | 94.0                 | 0.0           | $\pm 1.5$             |
| 114.0               | 114.0                | +0.1          | $\pm 1.5$             |

### REMARK:

1. The indication value was obtained from the average of ten replicated measurement.
2. The equipment being used in this calibration are regularly calibrated by laboratory according to ISO/IEC 17025.

-----End of Report-----

*PREPARED AND CHECKED BY:*

For and On Behalf of **High Precision Chemical Testing Limited**

*Laboratory Director (CHAN Hon-Fai)*

# High-Volume TSP Sampler

## 5-POINT CALIBRATION DATA SHEET



File No. MA16034/05/0031

Project No. AM1 - Tin Hau Temple

Date: 10-Aug-21 Next Due Date: 10-Oct-21 Operator: SK

Equipment No.: A-01-05 Model No.: GS2310 Serial No. 10599

| Ambient Condition   |            |                     |              |
|---------------------|------------|---------------------|--------------|
| Temperature, Ta (K) | <u>302</u> | Pressure, Pa (mmHg) | <u>754.3</u> |

| Orifice Transfer Standard Information |                  |  |                |               |                 |
|---------------------------------------|------------------|--|----------------|---------------|-----------------|
| Serial No.                            | <u>3864</u>      | Slope, mc  | <u>0.05846</u> | Intercept, bc | <u>-0.00313</u> |
| Last Calibration Date:                | <u>11-Jan-21</u> | $mc \times Qstd + bc = [\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ |                |               |                 |
| Next Calibration Date:                | <u>11-Jan-22</u> | $Qstd = \{[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |                |               |                 |

| Calibration of TSP Sampler |                                    |  |                        |                                |  |
|----------------------------|------------------------------------|--|------------------------|--------------------------------|--|
| Calibration Point          | Orifice                            |  |                        | HVS                            |  |
|                            | $\Delta H$ (orifice), in. of water | $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (CFM)<br>X - axis | $\Delta W$ (HVS), in. of water | $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$<br>Y-axis |
| 1                          | <u>13.2</u>                        | <u>3.60</u>  | <u>61.56</u>           | <u>9.6</u>                     | <u>3.07</u>  |
| 2                          | <u>9.6</u>                         | <u>3.07</u>  | <u>52.50</u>           | <u>7.4</u>                     | <u>2.69</u>  |
| 3                          | <u>7.4</u>                         | <u>2.69</u>  | <u>46.10</u>           | <u>5.4</u>                     | <u>2.30</u>  |
| 4                          | <u>5.2</u>                         | <u>2.26</u>  | <u>38.66</u>           | <u>3.4</u>                     | <u>1.82</u>  |
| 5                          | <u>3.0</u>                         | <u>1.71</u>  | <u>29.37</u>           | <u>2.0</u>                     | <u>1.40</u>  |

### By Linear Regression of Y on X

Slope, mw = 0.0535 Intercept, bw = -0.1854

Correlation coefficient\* = 0.9971

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation   |  |
|---|--|
| From the TSP Field Calibration Curve, take Qstd = 43 CFM  |  |
| From the Regression Equation, the "Y" value according to  |  |
| $mw \times Qstd + bw = [\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$                              |  |
| Therefore, Set Point; W = $(mw \times Qstd + bw)^2 \times (760 / Pa) \times (Ta / 298) =$ <u>4.57</u> |  |

Remarks: \_\_\_\_\_

Conducted by: Wong Shing Kwai Signature: [Signature] Date: 10-Aug-21

Checked by: Henry Leung Signature: [Signature] Date: 10-Aug-21



# High-Volume TSP Sampler

## 5-POINT CALIBRATION DATA SHEET



File No. MA16034/08/0031

Project No. AM2 - Sai Tso Wan Recreation Ground  
 Date: 10-Aug-21 Next Due Date: 10-Oct-21 Operator: SK  
 Equipment No.: A-01-08 Model No.: GS2310 Serial No. 1287

| Ambient Condition   |            |                     |              |
|---------------------|------------|---------------------|--------------|
| Temperature, Ta (K) | <u>302</u> | Pressure, Pa (mmHg) | <u>754.3</u> |

| Orifice Transfer Standard Information |                  |  |                |               |                 |
|---------------------------------------|------------------|--|----------------|---------------|-----------------|
| Serial No.                            | <u>3864</u>      | Slope, mc  | <u>0.05846</u> | Intercept, bc | <u>-0.00313</u> |
| Last Calibration Date:                | <u>11-Jan-21</u> | $mc \times Qstd + bc = [\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ |                |               |                 |
| Next Calibration Date:                | <u>11-Jan-22</u> | $Qstd = \{[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |                |               |                 |

| Calibration of TSP Sampler |                                    |  |                        |                                |  |
|----------------------------|------------------------------------|--|------------------------|--------------------------------|--|
| Calibration Point          | Orifice                            |  |                        | HVS                            |  |
|                            | $\Delta H$ (orifice), in. of water | $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (CFM)<br>X - axis | $\Delta W$ (HVS), in. of water | $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$<br>Y-axis |
| 1                          | <u>13.4</u>                        | <u>3.62</u>  | <u>62.02</u>           | <u>9.0</u>                     | <u>2.97</u>  |
| 2                          | <u>10.2</u>                        | <u>3.16</u>  | <u>54.12</u>           | <u>6.4</u>                     | <u>2.50</u>  |
| 3                          | <u>7.9</u>                         | <u>2.78</u>  | <u>47.63</u>           | <u>4.9</u>                     | <u>2.19</u>  |
| 4                          | <u>5.1</u>                         | <u>2.23</u>  | <u>38.28</u>           | <u>3.3</u>                     | <u>1.80</u>  |
| 5                          | <u>3.0</u>                         | <u>1.71</u>  | <u>29.37</u>           | <u>2.0</u>                     | <u>1.40</u>  |



### By Linear Regression of Y on X

Slope, mw = 0.0472 Intercept, bw = -0.0147  
 Correlation coefficient\* = 0.9976

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation   |  |
|---|--|
| From the TSP Field Calibration Curve, take Qstd = 43 CFM  |  |
| From the Regression Equation, the "Y" value according to  |  |
| $mw \times Qstd + bw = [\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$                              |  |
| Therefore, Set Point; W = $(mw \times Qstd + bw)^2 \times (760 / Pa) \times (Ta / 298) =$ <u>4.15</u> |  |

Remarks: \_\_\_\_\_

Conducted by: Wong Shing Kwai Signature:  Date: 10-Aug-21  
 Checked by: Henry Leung Signature:  Date: 10-Aug-21

# High-Volume TSP Sampler

## 5-POINT CALIBRATION DATA SHEET



File No. MA16034/03/0031

Project No. AM3 - Yau Lai Estate, Bik Lai House

Date: 10-Aug-21 Next Due Date: 10-Oct-21 Operator: SK

Equipment No.: A-01-03 Model No.: GS2310 Serial No. 10379

| Ambient Condition   |            |                     |              |
|---------------------|------------|---------------------|--------------|
| Temperature, Ta (K) | <u>302</u> | Pressure, Pa (mmHg) | <u>754.3</u> |

| Orifice Transfer Standard Information |                  |  |                |               |                 |
|---------------------------------------|------------------|--|----------------|---------------|-----------------|
| Serial No.                            | <u>3864</u>      | Slope, mc  | <u>0.05846</u> | Intercept, bc | <u>-0.00313</u> |
| Last Calibration Date:                | <u>11-Jan-21</u> | $mc \times Qstd + bc = [\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ |                |               |                 |
| Next Calibration Date:                | <u>11-Jan-22</u> | $Qstd = \{[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |                |               |                 |

| Calibration of TSP Sampler |                                    |  |                        |                                |  |
|----------------------------|------------------------------------|--|------------------------|--------------------------------|--|
| Calibration Point          | Orifice                            |  |                        | HVS                            |  |
|                            | $\Delta H$ (orifice), in. of water | $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (CFM)<br>X - axis | $\Delta W$ (HVS), in. of water | $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$<br>Y-axis |
| 1                          | <u>13.2</u>                        | <u>3.60</u>  | <u>61.56</u>           | <u>9.0</u>                     | <u>2.97</u>  |
| 2                          | <u>10.2</u>                        | <u>3.16</u>  | <u>54.12</u>           | <u>6.8</u>                     | <u>2.58</u>  |
| 3                          | <u>8.0</u>                         | <u>2.80</u>  | <u>47.93</u>           | <u>5.4</u>                     | <u>2.30</u>  |
| 4                          | <u>5.4</u>                         | <u>2.30</u>  | <u>39.39</u>           | <u>3.5</u>                     | <u>1.85</u>  |
| 5                          | <u>2.9</u>                         | <u>1.69</u>  | <u>28.88</u>           | <u>2.0</u>                     | <u>1.39</u>  |

### By Linear Regression of Y on X

Slope, mw = 0.0484 Intercept, bw = -0.0251

Correlation coefficient\* = 0.9994

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation   |  |
|---|--|
| From the TSP Field Calibration Curve, take Qstd = 43 CFM  |  |
| From the Regression Equation, the "Y" value according to  |  |
| $mw \times Qstd + bw = [\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$                              |  |
| Therefore, Set Point; W = $(mw \times Qstd + bw)^2 \times (760 / Pa) \times (Ta / 298) =$ <u>4.31</u> |  |

Remarks: \_\_\_\_\_

Conducted by: Wong Shing Kwai Signature: [Signature] Date: 10-Aug-21

Checked by: Henry Leung Signature: [Signature] Date: 10-Aug-21

# High-Volume TSP Sampler

## 5-POINT CALIBRATION DATA SHEET



File No. MA16034/54/0031

Project No. AM4(A) - Cha Kwo Ling Public Cargo Working Area Administrative Office

Date: 10-Aug-21 Next Due Date: 10-Oct-21 Operator: SK

Equipment No.: A-01-54 Model No.: TE-5170 Serial No. 1536

| Ambient Condition   |            |                     |              |
|---------------------|------------|---------------------|--------------|
| Temperature, Ta (K) | <u>302</u> | Pressure, Pa (mmHg) | <u>754.3</u> |

| Orifice Transfer Standard Information |                  |  |                |               |                 |
|---------------------------------------|------------------|--|----------------|---------------|-----------------|
| Serial No.                            | <u>3864</u>      | Slope, mc  | <u>0.05846</u> | Intercept, bc | <u>-0.00313</u> |
| Last Calibration Date:                | <u>11-Jan-21</u> | $mc \times Qstd + bc = [\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ |                |               |                 |
| Next Calibration Date:                | <u>11-Jan-22</u> | $Qstd = \{[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |                |               |                 |

| Calibration of TSP Sampler |                                    |  |                        |                                |  |
|----------------------------|------------------------------------|--|------------------------|--------------------------------|--|
| Calibration Point          | Orifice                            |  |                        | HVS                            |  |
|                            | $\Delta H$ (orifice), in. of water | $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (CFM)<br>X - axis | $\Delta W$ (HVS), in. of water | $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$<br>Y-axis |
| 1                          | <u>13.4</u>                        | <u>3.62</u>  | <u>62.02</u>           | <u>9.4</u>                     | <u>3.03</u>  |
| 2                          | <u>10.8</u>                        | <u>3.25</u>  | <u>55.69</u>           | <u>7.2</u>                     | <u>2.66</u>  |
| 3                          | <u>7.6</u>                         | <u>2.73</u>  | <u>46.72</u>           | <u>5.1</u>                     | <u>2.23</u>  |
| 4                          | <u>5.6</u>                         | <u>2.34</u>  | <u>40.11</u>           | <u>3.6</u>                     | <u>1.88</u>  |
| 5                          | <u>3.0</u>                         | <u>1.71</u>  | <u>29.37</u>           | <u>1.9</u>                     | <u>1.36</u>  |

### By Linear Regression of Y on X


Slope, mw = 0.0508 Intercept, bw = -0.1424


Correlation coefficient\* = 0.9994

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation   |  |
|---|--|
| From the TSP Field Calibration Curve, take Qstd = 43 CFM  |  |
| From the Regression Equation, the "Y" value according to  |  |
| $mw \times Qstd + bw = [\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$                              |  |
| Therefore, Set Point; W = $(mw \times Qstd + bw)^2 \times (760 / Pa) \times (Ta / 298) =$ <u>4.25</u> |  |

Remarks: \_\_\_\_\_

Conducted by: Wong Shing Kwai Signature:  Date: 10-Aug-21

Checked by: Henry Leung Signature:  Date: 10-Aug-21

# High-Volume TSP Sampler

## 5-POINT CALIBRATION DATA SHEET



File No. MA16034/05/0032

Project No. AM1 - Tin Hau Temple

Date: 9-Oct-21 Next Due Date: 9-Dec-21 Operator: SK

Equipment No.: A-01-05 Model No.: GS2310 Serial No. 10599

| Ambient Condition   |              |                     |              |
|---------------------|--------------|---------------------|--------------|
| Temperature, Ta (K) | <u>299.5</u> | Pressure, Pa (mmHg) | <u>753.6</u> |

| Orifice Transfer Standard Information |                  |  |                |               |                 |
|---------------------------------------|------------------|--|----------------|---------------|-----------------|
| Serial No.                            | <u>3864</u>      | Slope, mc  | <u>0.05846</u> | Intercept, bc | <u>-0.00313</u> |
| Last Calibration Date:                | <u>11-Jan-21</u> | $mc \times Qstd + bc = [\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ |                |               |                 |
| Next Calibration Date:                | <u>11-Jan-22</u> | $Qstd = \{[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |                |               |                 |

| Calibration of TSP Sampler |                                    |  |                        |                                |  |
|----------------------------|------------------------------------|--|------------------------|--------------------------------|--|
| Calibration Point          | Orifice                            |  |                        | HVS                            |  |
|                            | $\Delta H$ (orifice), in. of water | $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (CFM)<br>X - axis | $\Delta W$ (HVS), in. of water | $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$<br>Y-axis |
| 1                          | <u>13.3</u>                        | <u>3.62</u>  | <u>62.02</u>           | <u>9.1</u>                     | <u>3.00</u>  |
| 2                          | <u>9.8</u>                         | <u>3.11</u>  | <u>53.24</u>           | <u>7.0</u>                     | <u>2.63</u>  |
| 3                          | <u>7.4</u>                         | <u>2.70</u>  | <u>46.27</u>           | <u>5.2</u>                     | <u>2.27</u>  |
| 4                          | <u>5.2</u>                         | <u>2.27</u>  | <u>38.80</u>           | <u>3.2</u>                     | <u>1.78</u>  |
| 5                          | <u>3.0</u>                         | <u>1.72</u>  | <u>29.48</u>           | <u>2.0</u>                     | <u>1.40</u>  |

### By Linear Regression of Y on X

Slope, mw = 0.0506 Intercept, bw = -0.1108

Correlation coefficient\* = 0.9968

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation   |  |
|---|--|
| From the TSP Field Calibration Curve, take Qstd = 43 CFM  |  |
| From the Regression Equation, the "Y" value according to  |  |
| $mw \times Qstd + bw = [\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$                              |  |
| Therefore, Set Point; W = $(mw \times Qstd + bw)^2 \times (760 / Pa) \times (Ta / 298) =$ <u>4.32</u> |  |

Remarks: \_\_\_\_\_

Conducted by: Wong Shing Kwai Signature: [Signature] Date: 9-Oct-21

Checked by: Henry Leung Signature: [Signature] Date: 9-Oct-21

# High-Volume TSP Sampler

## 5-POINT CALIBRATION DATA SHEET



File No. MA16034/08/0032

Project No. AM2 - Sai Tso Wan Recreation Ground  
 Date: 9-Oct-21 Next Due Date: 9-Dec-21 Operator: SK  
 Equipment No.: A-01-08 Model No.: GS2310 Serial No. 1287

| Ambient Condition   |              |                     |              |
|---------------------|--------------|---------------------|--------------|
| Temperature, Ta (K) | <u>299.5</u> | Pressure, Pa (mmHg) | <u>753.6</u> |

| Orifice Transfer Standard Information |                  |  |                |               |                 |
|---------------------------------------|------------------|--|----------------|---------------|-----------------|
| Serial No.                            | <u>3864</u>      | Slope, mc  | <u>0.05846</u> | Intercept, bc | <u>-0.00313</u> |
| Last Calibration Date:                | <u>11-Jan-21</u> | $mc \times Qstd + bc = [\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ |                |               |                 |
| Next Calibration Date:                | <u>11-Jan-22</u> | $Qstd = \{[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |                |               |                 |

| Calibration of TSP Sampler |                                    |  |                        |                                |  |
|----------------------------|------------------------------------|--|------------------------|--------------------------------|--|
| Calibration Point          | Orifice                            |  |                        | HVS                            |  |
|                            | $\Delta H$ (orifice), in. of water | $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (CFM)<br>X - axis | $\Delta W$ (HVS), in. of water | $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$<br>Y-axis |
| 1                          | <u>13.2</u>                        | <u>3.61</u>  | <u>61.78</u>           | <u>9.0</u>                     | <u>2.98</u>  |
| 2                          | <u>10.2</u>                        | <u>3.17</u>  | <u>54.32</u>           | <u>6.6</u>                     | <u>2.55</u>  |
| 3                          | <u>7.9</u>                         | <u>2.79</u>  | <u>47.81</u>           | <u>4.9</u>                     | <u>2.20</u>  |
| 4                          | <u>5.2</u>                         | <u>2.27</u>  | <u>38.80</u>           | <u>3.3</u>                     | <u>1.80</u>  |
| 5                          | <u>3.0</u>                         | <u>1.72</u>  | <u>29.48</u>           | <u>2.0</u>                     | <u>1.40</u>  |



### By Linear Regression of Y on X

Slope, mw = 0.0484 Intercept, bw = -0.0605  
 Correlation coefficient\* = 0.9976

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation   |  |
|---|--|
| From the TSP Field Calibration Curve, take Qstd = 43 CFM  |  |
| From the Regression Equation, the "Y" value according to  |  |
| $mw \times Qstd + bw = [\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$                              |  |
| Therefore, Set Point; W = $(mw \times Qstd + bw)^2 \times (760 / Pa) \times (Ta / 298) =$ <u>4.14</u> |  |

Remarks: \_\_\_\_\_

Conducted by: Wong Shing Kwai Signature:  Date: 9-Oct-21  
 Checked by: Henry Leung Signature:  Date: 9-Oct-21

# High-Volume TSP Sampler

## 5-POINT CALIBRATION DATA SHEET



File No. MA16034/03/0032

Project No. AM3 - Yau Lai Estate, Bik Lai House

Date: 9-Oct-21

Next Due Date: 9-Dec-21

Operator: SK

Equipment No.: A-01-03

Model No.: GS2310

Serial No. 10379

| Ambient Condition   |              |                     |              |
|---------------------|--------------|---------------------|--------------|
| Temperature, Ta (K) | <u>299.5</u> | Pressure, Pa (mmHg) | <u>753.6</u> |

| Orifice Transfer Standard Information |                  |  |                |               |                 |
|---------------------------------------|------------------|--|----------------|---------------|-----------------|
| Serial No.                            | <u>3864</u>      | Slope, mc  | <u>0.05846</u> | Intercept, bc | <u>-0.00313</u> |
| Last Calibration Date:                | <u>11-Jan-21</u> | $mc \times Qstd + bc = [\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ |                |               |                 |
| Next Calibration Date:                | <u>11-Jan-22</u> | $Qstd = \{[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |                |               |                 |

| Calibration of TSP Sampler |                                    |  |                        |                                |  |
|----------------------------|------------------------------------|--|------------------------|--------------------------------|--|
| Calibration Point          | Orifice                            |  |                        | HVS                            |  |
|                            | $\Delta H$ (orifice), in. of water | $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (CFM)<br>X - axis | $\Delta W$ (HVS), in. of water | $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$<br>Y-axis |
| 1                          | <u>13.3</u>                        | <u>3.62</u>  | <u>62.02</u>           | <u>9.1</u>                     | <u>3.00</u>  |
| 2                          | <u>10.3</u>                        | <u>3.19</u>  | <u>54.58</u>           | <u>6.8</u>                     | <u>2.59</u>  |
| 3                          | <u>8.2</u>                         | <u>2.84</u>  | <u>48.71</u>           | <u>5.4</u>                     | <u>2.31</u>  |
| 4                          | <u>5.6</u>                         | <u>2.35</u>  | <u>40.26</u>           | <u>3.5</u>                     | <u>1.86</u>  |
| 5                          | <u>2.9</u>                         | <u>1.69</u>  | <u>28.99</u>           | <u>2.0</u>                     | <u>1.40</u>  |

### By Linear Regression of Y on X

Slope, mw = 0.0486

Intercept, bw = -0.0498

Correlation coefficient\* = 0.9983

\*If Correlation Coefficient < 0.990, check and recalibrate.

### Set Point Calculation

From the TSP Field Calibration Curve, take Qstd = 43 CFM

From the Regression Equation, the "Y" value according to

$$mw \times Qstd + bw = [\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$$

Therefore, Set Point; W =  $(mw \times Qstd + bw)^2 \times (760 / Pa) \times (Ta / 298) =$  4.22

Remarks:

Conducted by: Wong Shing Kwai

Signature: 

Date: 9-Oct-21

Checked by: Henry Leung

Signature: 

Date: 9-Oct-21

# High-Volume TSP Sampler

## 5-POINT CALIBRATION DATA SHEET



File No. MA16034/54/0032

Project No. AM4(A) - Cha Kwo Ling Public Cargo Working Area Administrative Office

Date: 9-Oct-21 Next Due Date: 9-Dec-21 Operator: SK

Equipment No.: A-01-54 Model No.: TE-5170 Serial No. 1536

| Ambient Condition   |              |                     |              |
|---------------------|--------------|---------------------|--------------|
| Temperature, Ta (K) | <u>299.5</u> | Pressure, Pa (mmHg) | <u>753.6</u> |

| Orifice Transfer Standard Information |                  |  |                |               |                 |
|---------------------------------------|------------------|--|----------------|---------------|-----------------|
| Serial No.                            | <u>3864</u>      | Slope, mc  | <u>0.05846</u> | Intercept, bc | <u>-0.00313</u> |
| Last Calibration Date:                | <u>11-Jan-21</u> | $mc \times Qstd + bc = [\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ |                |               |                 |
| Next Calibration Date:                | <u>11-Jan-22</u> | $Qstd = \{[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2} - bc\} / mc$  |                |               |                 |

| Calibration of TSP Sampler |                                    |  |                        |                                |  |
|----------------------------|------------------------------------|--|------------------------|--------------------------------|--|
| Calibration Point          | Orifice                            |  |                        | HVS                            |  |
|                            | $\Delta H$ (orifice), in. of water | $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ | Qstd (CFM)<br>X - axis | $\Delta W$ (HVS), in. of water | $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$<br>Y-axis |
| 1                          | <u>13.2</u>                        | <u>3.61</u>  | <u>61.78</u>           | <u>9.4</u>                     | <u>3.05</u>  |
| 2                          | <u>10.6</u>                        | <u>3.23</u>  | <u>55.37</u>           | <u>7.4</u>                     | <u>2.70</u>  |
| 3                          | <u>7.6</u>                         | <u>2.74</u>  | <u>46.89</u>           | <u>5.1</u>                     | <u>2.24</u>  |
| 4                          | <u>5.6</u>                         | <u>2.35</u>  | <u>40.26</u>           | <u>3.6</u>                     | <u>1.88</u>  |
| 5                          | <u>3.0</u>                         | <u>1.72</u>  | <u>29.48</u>           | <u>1.9</u>                     | <u>1.37</u>  |

### By Linear Regression of Y on X


Slope, mw = 0.0522 Intercept, bw = -0.1914


Correlation coefficient\* = 0.9996

\*If Correlation Coefficient < 0.990, check and recalibrate.

| Set Point Calculation   |  |
|---|--|
| From the TSP Field Calibration Curve, take Qstd = 43 CFM  |  |
| From the Regression Equation, the "Y" value according to  |  |
| $mw \times Qstd + bw = [\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$                              |  |
| Therefore, Set Point; W = $(mw \times Qstd + bw)^2 \times (760 / Pa) \times (Ta / 298) =$ <u>4.27</u> |  |

Remarks: \_\_\_\_\_

Conducted by: Wong Shing Kwai Signature:  Date: 9-Oct-21

Checked by: Henry Leung Signature:  Date: 9-Oct-21



## Calibration Certificate

0025914

|   |   |
|---|---|
| Customer :<br>Cinotech Consultants Limited<br>RM 1710, Technology Park,<br>18 On Lai Street, Shatin, N.T.<br>Hong Kong<br><br>Customer Code : SVEC09005 | Object 1 : SVAN957 SLM<br>Serial No. /Ref. No. : 23851 / N-08-12<br>Object 2 : Microphone<br>Serial No. /Ref. No. : 43676<br><br>Manufacturer : Svantek |
| Date of calibration: 22/01/2021<br>Date of the recommended re-calibration: 22/01/2022   | Certificate No.: 0025914<br>Handle by: E0002  |

### Measuring results

| Reference value | Indication value | Deviation | Allowed deviation | Object |
|-----------------|------------------|-----------|-------------------|--------|
| 94.0dB          | 93.6dB           | -0.4dB    | +/- 1.5dB         | 1      |
| 114.0dB         | 113.5dB          | -0.5dB    | +/- 1.5dB         | 1      |

### Measuring equipment

| index | Calibrator / Master                 | Traceability |
|-------|-------------------------------------|--------------|
| 1     | Master Sound Meter, SVAN949,sn:8571 | IEC61672     |
| 2     | Sound Calibrator, SV30A sn:32580    | IEC60942     |

### Ambient conditions

Temperature (20...26)°C

Humidity (20...60)%RH

### Measuring procedure

Calibrated by Type 1 Sound Calibrator with Master Sound Level Meter under 1kHz Frequency.

### Uncertainty

+/- 0.2 dB for probability not less than 95%.

### Conformity

- 1.The resulted values were those obtained at the time of test and applies only to the item calibrated.
- 2.The measurement uncertainty was calculated according to the regulations of GUM with the coverage factor k=2 and contains the uncertainty of the measuring procedure and the uncertainty of the measuring system.
- 3.The equipment being used in this calibration are regularly calibrated by laboratory according to ISO/IEC17025.
- 4.HKAS has accredited this laboratory (HOKLAS 267) for specific calibration activities as listed in the HOKLAS directory of accredited laboratories.
- 5.The calibrations certificate may not be reproduced.

Measured value(s) **within** the allowable deviation.

Performed by

Calibration Technician

Approved by

Quality Manager





Equipment no.: N-13-01

**Calibration Certificate****0025247**

|  |  |
|--|--|
| Customer :<br>Cinotech Consultants Limited<br>RM 1710, Technology Park,<br>18 On Lai Street, Shatin, N.T.<br>Hong Kong | Object 1 : ST-120 sound calibrator<br>Serial No. /Ref. No. : 181001608<br>Object 2 :<br>Serial No. /Ref. No. : |
| Customer Code : SVEC09005  | Manufacturer : Soundtek  |
| Date of calibration: 05/11/2020<br>Date of the recommended re-calibration: 05/11/2021                                  | Certificate No.: 0025247<br>Handle by: E0002   |

**Measuring results**

| Reference value | Indication value | Deviation | Allowed deviation | Object |
|-----------------|------------------|-----------|-------------------|--------|
| 94.0dB          | 93.7dB           | -0.3dB    | +/- 0.3dB         | 1      |
| 114.0dB         | 113.6dB          | -0.4dB    | +/- 0.5dB         | 1      |

**Measuring equipment**

| index | Calibrator / Master                 | Traceability |
|-------|-------------------------------------|--------------|
| 1     | Master Sound Meter, SVAN949,sn:8571 | IEC61672     |
| 2     | Sound Calibrator, SV30A sn:32580    | IEC60942     |

**Ambient conditions**

Temperature (20...26)°C

Humidity (20...60)%RH

**Measuring procedure**

Calibrated by Type 1 Sound Level Meter and 1kHz Sound Source

**Uncertainty**

+/- 0.2 dB for probability not less than 95%.

**Conformity**

- 1.The resulted values were those obtained at the time of test and applies only to the item calibrated.
- 2.The measurement uncertainty was calculated according to the regulations of GUM with the coverage factor k=2 and contains the uncertainty of the measuring procedure and the uncertainty of the measuring system.
- 3.The equipment being used in this calibration are regularly calibrated by laboratory according to ISO/IEC17025.
- 4.HKAS has accredited this laboratory (HOKLAS 267) for specific calibration activities as listed in the HOKLAS directory of accredited laboratories.
- 5.The calibrations certificate may not be reproduced.

**Measured value(s)** **within** **the allowable deviation.**

Performed by

Calibration Technician

Mr. K.L. Ng

Approved by

Quality Manager

**Certificate of Calibration**

It is certified that the item under calibration has been calibrated by corresponding calibrated High Volume Sampler


Description: Digital Dust Indicator Date of Calibration 2-Aug-21  
 Manufacturer: Sibata Scientific Technology LTD. Validity of Calibration Record 2-Oct-21  
 Model No.: LD-5R  
 Serial No.: 972778  
 Equipment No.: SA-01-07 Sensitivity 0.001 mg/m3  
 High Volume Sampler No.: A-01-03 Before Sensitivity Adjustment 735 CPM  
 Tisch Calibration Orifice No.: 3864 After Sensitivity Adjustment 735 CPM


| Calibration of 1 hr TSP   |   |   |
|---|---|---|
| Calibration Point   | Laser Dust Monitor  | HVS   |
|   | Mass Concentration ( $\mu\text{g}/\text{m}^3$ )<br>X-axis | Mass concentration ( $\mu\text{g}/\text{m}^3$ )<br>Y-axis |
| 1   | 61.0  | 131.0   |
| 2   | 56.0  | 125.0   |
| 3   | 48.0  | 116.0   |
| Average   | 55.0  | 124.0   |
| By Linear Regression of Y on X<br>Slope , mw = <u>1.1512</u> Intercept, bw = <u>60.6860</u><br>Correlation coefficient* = <u>0.9998</u> |   |   |
| Set Correlation Factor  |   |   |
| Particulate Concentration by High Volume Sampler ( $\mu\text{g}/\text{m}^3$ )   |   | 124.0   |
| Particulate Concentration by Dust Meter ( $\mu\text{g}/\text{m}^3$ )  |   | 55.0  |
| Measureing time, (min)  |   | 60.0  |
| Set Correlation Factor , SCF  |   |   |
| SCF = [ K=High Volume Sampler / Dust Meter, ( $\mu\text{g}/\text{m}^3$ ) ]  |   | <u>2.3</u>  |

In-house method in according to the instruction manual:

The Dust Monitor was compared with a calibrated High Volume Sampler and The result was used to generate the Correlation Factor (CF) between the Dust Monitor and High Volume Sampler.

Those filter papers are weighted by HOKLAS laboratory (HPCT Litimed)

Calibrated by:   
 Technical Officer (Wong Shing Kwai)

Approved by:   
 Project Manager (Henry Leung)